

## User Manual

### Audiolyser® ADL-20



# Table of Contents

<b>1. Introduction</b>	<b>4</b>
1.1. List of equipment supplied	5
1.2. Audiometer presentation	6
1.3. Technical features	7
1.4. Multilingual vocal assistant (MVA)	8
1.5. Symbols	8
<b>2. Safety</b>	<b>9</b>
2.1. General	10
2.1. Users	10
2.2. Medical contraindications	11
2.3. Environment for use	11
<b>3. Installation</b>	<b>12</b>
3.1. Prerequisite	13
3.1.1. Software recommendations	13
3.1.2. Equipment recommendations	13
3.2. Software use conditions	13
3.3. Installation procedure	14
3.4. Stop the device	17
<b>4. Use</b>	<b>18</b>
4.1. Using the hygienic single-use earpad covers	19
4.2. Preliminary patient explanation	19
4.2.1. Headset position	19
4.2.2. MVA (multilingual vocal assistant)	19
4.2.3. Most frequent errors	20
4.3. Software use	21
4.3.1. Presentation	21
4.3.2. Checking function	21
4.3.3. Tool bar	22
4.3.4. Menu bar	22
4.4. Patient identification	23
4.4.1. Identification	23
4.4.2. Display of predicted values	24
4.5. Open a patient file/test	24

4.5.1.	Open a file	24
4.5.2.	Open a test	25
<b>4.6.</b>	<b>Parameter the software</b>	<b>25</b>
4.6.1.	Screening	25
4.6.2.	Window parameters	26
<b>4.7.</b>	<b>Display areas categorization</b>	<b>31</b>
<b>4.8.</b>	<b>Manual mode</b>	<b>33</b>
4.8.1.	Operation	33
4.8.2.	Keyboard keys	34
<b>4.9.</b>	<b>Automatic mode</b>	<b>35</b>
4.9.1.	Parameters	35
4.9.2.	Creation of sequences	36
4.9.3.	Run automatic test	37
<b>4.10.</b>	<b>Display of results</b>	<b>37</b>
<b>4.11.</b>	<b>Print results</b>	<b>38</b>
4.11.1.	Simple printing	38
4.11.2.	Selective printing	39
4.11.3.	PDF printing	39
<b>5.</b>	<b>Maintenance</b>	<b>41</b>
5.1.	Cleaning	42
5.2.	Maintenance	42
5.3.	Guarantee	42
<b>6.</b>	<b>Available Accessories</b>	<b>43</b>
6.1.	Hygienic single-use earpad covers	44
<b>7.</b>	<b>FAQ</b>	<b>45</b>
7.1.	No sound is perceived	46
7.2.	Error message appears at recording	46
7.2.1.	“Identification incomplete”	46
7.2.2.	“Operation must use updateable query”	46
7.3.	Patient identification file found but not the tests	46

## 1. Introduction

## 1.1. List of equipment supplied

---

The following equipment should be present in the packaging:

- AUDIOLYSER® ADL-20 device with headset and USB lead (in the cover)
- CD Rom containing user manual and Audiowin®20 software
- Information sheet
- Calibration certificate
- Transport cover


**Note:** The AUDIOLYSER® ADL-20 should be returned to the After Sales Service (for calibration or other services) in its original packaging and padding. Guarantee repairs will not be accepted without the original packaging box.

## 1.2. Audiometer presentation

The AUDIOLYSER® ADL-20 is a computerized digital audiometer. The electronics are integrated into the patient response button, rendering the device lightweight and easy to transport.

The AUDIOLYSER® ADL-20 can be set up with different headsets depending on user needs (refer §1.3).

One of the innovations of the AUDIOLYSER® ADL-20 is the Multilingual Vocal Assistant (MVA) which gives the patient the necessary instructions for the test and signals, if necessary, incorrect patient handling.



The DSP (Digital Signal Processor) located in the patient response button ensures communication with both the computer and the sound generation.

The AUDIOLYSER® ADL-20 is driven by the Audiowin® 20 software, a simple and intuitive interface. Audiowin® 20 stores information in a data base, which can also be printed, recorded and exported to other software.

Storage of audiometric curves and results enables consultation of files at a later date as well as statistics processing of results.

With all its features, the AUDIOLYSER® ADL-20 is a reliable, high-performance and progressive tool.

## 1.3. Technical features

AUDIOLYSER® ADL-20 features:	
Modes of use	Manual or automatic
Sound transmission	Continuous, inversed or pulsed mode
Harmonic distortion	± 2.5%
Frequency accuracy	± 2%
Headset soundproofing	10 - 40 dB (depending on model)
Length of lead	3 metres
Storage temperature	0 - 50°C
Temperature for use	15 - 35°C
Humidity	75% maximum
Altitude of operation	< 2000 metres
Voltage	5VDC (via USB port)
Supply current	210mA maximum
Reference standards	EN 60601-1, EN 60601-1-2, EN 60645-1, ISO 8253-1, ISO 389-1, ISO 389-5, ISO 389-8, ISO 7029
Type of audiometer	4
Medical class	Ila
Applied part	BF Type
Dimensions in packaging	255 x 210 x 100 mm
Total weight of device	500 - 850g approximately (depending on model)

Specific features of each headset:					
Type	BEYER DYNAMIC DT 48	TELEPHONICS TDH 39	SENNHEISER HDA 200	SENNHEISER HDA 300	HOLMCO PD-81
Calibration standard	ISO 389-1	ISO 389-1	ISO 389-5 ISO 389-8	ISO 389-5 ISO 389-8	Manufacturer
Headset weight	600g	430g	445g	490g	725g

Intensity limit for each headset:											
Frequencies (Hz)	125	250	500	750	1000	1500	2000	3000	4000	6000	8000
TELEPHONICS TDH 39 (dB)	70	90	100	110	110	110	110	110	100	90	80
SENNHEISER HDA 200 (dB)	70	90	100	100	100	100	100	100	90	90	80
SENNHEISER HDA 300	70	90	100	110	110	110	110	110	110	110	100
BEYER DYNAMIC DT 48 (dB)	70	90	100	110	110	110	110	110	110	90	80
HOLMCO PD-81 (dB)	70	80	90	100	100	100	100	100	100	90	80

## 1.4. Multilingual vocal assistant (MVA)

The AUDIOLYSER® ADL-20 has integrated a multilingual vocal assistant (MVA) system. By default, only the language of your country is available, but other languages may also be purchased (**French, English, Arab, German, Chinese Mandarin, Dutch**).

## 1.5. Symbols

Serial number labels display the following information:

	CE Marking Directive 93/42/EEC + N° of notified body
0459	
	BF type applied part
	Device should not be discarded with general household waste. Return it to the manufacturer for disposal
	Consult the accompanying documents
SN	Serial number
	Manufacturer identification
	Single use
(01)XXXXXXXXXXXXXXXXXX	UDI Identification
(10)XXXXXX	UDI Identification Batch number
(11)XXXXXX	UDI Identification Date of manufacture
(17)XXXXXX	UDI Identification Expiry date

## 2. Safety

## 2.1. General

---

### **Caution:**

- Do not use the AUDIOLYSER® ADL-20 in a non-medical environment.
- Do not expose a person to levels above 100 dB without having read the safety instructions (or without authorization).
- Do not dismantle or interfere with the internal components.
- Do not plug in or use the AUDIOLYSER® ADL-20 in an explosive environment or in the presence of anaesthetic gases.
- Do not drop the headset or the device.
- Do not splash or wet the headphones, even slightly (be careful of sprays and aerosol disinfectants).
- Clean only the headset pads (refer §5.2).
- Do not pull on the headset leads.
- Do not make sudden movements while wearing the headset.
- Do not let the PC and AUDIOLYSER® ADL-20 leads dangle on the ground and become entwined. This may cause it to fall, or degrade electrical connections.
- The operator should know whether the patient has an ear impairment and if a particular frequency or intensity could damage them. If this is the case, the operator should not perform the audiometry, or should be assisted by an authorized person who would be responsible for the test.
- The operator should ask the patient to remove glasses, hair accessories and/or hearing devices prior to an audiometry.
- Always use the AUDIOLYSER® ADL-20 connected to a computer according to IEC 60950-1
- In the conditions of use of the ambient test environment (see section 1.3), the AUDIOLYSER® ADL20 does not need warm-up period.
- For hygiene and biocompatibility reasons, it is compulsory to use the FIM MEDICAL brand of single-use earpad covers with audiometric headphones.  
These earpad covers have been developed specifically to meet ISO 10993 material biocompatibility requirements as well as guaranteeing perfect sound transmission whilst respecting IEC 60645-1.

**Caution: to meet ISO 10993 material biocompatibility requirements, it is compulsory to use the FIM MEDICAL brand of hygienic single-use earpad covers with Holmco PD-81, Sennheiser HDA200 and HDA300 audiometric headphones.**

## 2.1. Users

---

The AUDIOLYSER® ADL-20 should only be used by health professionals (doctor, hearing aid specialist, ENT specialist ...). For health professionals, the device should not prove difficult to use.

The operator should be sufficiently trained in the use of computers and the associated operating system. Curves and results should be interpreted by doctors having completed recognized training in the use of audiometers and the interpretation of audiograms. The operator should be aware of hygiene rules and bacterial contamination. If in doubt, the health professional should refer to this user manual and/or contact the FIM MEDICAL company or their distributor.

## 2.2. Medical contraindications

---

- Do not perform an audiometry on patients with auditory prostheses.
- Do not perform an audiometry without prior otoscopic examination.

## 2.3. Environment for use

---

A noisy environment can disturb audiometry tests and results. In this case, we recommend an audiometric booth.

## 3. Installation

## 3.1. Prerequisite

---

### 3.1.1. Software recommendations

Prerequisites application software Audiowin20<sup>®</sup> are:

- Windows XP, Windows Vista, Windows 7, Windows 8

### 3.1.2. Equipment recommendations

Prerequisites application equipment Audiowin20<sup>®</sup> are:

- PC with a 1 GHz processor at least
- 512 Mo of RAM
- 500 Mo free space on the hard disk
- A graphic card (or graphic chipset) accepting a resolution of at least 1024x748
- A keyboard
- A mouse
- A USB port
- A screen with a resolution accepting a minimum resolution of 1024x748

## 3.2. Software use conditions

---

The Audiowin<sup>®</sup> software is licenced under the conditions defined below. If, before installation, or 48 hours after receiving the software, these conditions are refused, please return it to the distributor to receive a refund.

Software use implies total agreement with the following conditions:

The software supplied under licence remains the property of FIM MEDICAL who grants the right to use this product as long as the present conditions are respected.

This licence is granted for installation and use at one workstation only (desk top computer, lap top or terminal).

All new installations require the purchase of a new licence or uninstallation from the initial workstation.

The licence for use is nominative and should not be passed on or sold without the written agreement of FIM MEDICAL.

Copy or reproduction of FIM MEDICAL software supplied under licence is prohibited. Reproduction, even partial, of original screens or computer processes constitutes a violation of this agreement. The user accepts to take the necessary measures to avoid pirate copies or use by non-authorized third parties.

The FIM MEDICAL company cannot be held responsible in any way for malfunctions related to the installation of one of their programmes on a computer. Neither can the FIM MEDICAL company be held responsible for any consequences related to the installation of one of their programmes, such as partial or total loss of data.


The user should be trained in the basic rules of handling computers. All precautions should be taken to guard against the risks of pirating programmes, divulging confidential data, attack by whatever type of computer virus or incorrect handling.

Special care should be taken to back up data recorded on the computer as often as possible, on a reliable media; we recommend performing this every day.

### 3.3. Installation procedure

The operator must possess all the administration rights on the workstation to be installed. Installation should be made in a clear space on a stable desk or table. The patient should not be able to see the screen, or the operator actions.


1. Switch on the computer.
2. Insert the Audiowin®20 software installation CD Rom into the CD reader.
3. Proceed to the software installation phase.
4. If your computer is set up for auto run, the installation procedure will run automatically.
5. If not, open Windows Explorer and search for "setup.exe" file in the root of the CD Rom, and run it.
6. Once the installation programme is run, follow the instructions.
7. Choose your installation language then click OK:




8. Click Next:




9. Enter the user code located on the back of the CD sleeve, then click Next:




10. Accept the agreement and click « Next » :




11. Choose the Audiowin® 20 software installation path (optional), then click Next:




12. Choose the data base installation path (*optional*), then click Next :




13. Click Next:



14. Click Install:



15. Wait while install :



16. Click Finish :



Locate a **USB** connector on the computer and plug in the ADL-20.

The USB lead can pass through the wall of a booth. Request information from your booth manufacturer or distributor.

### 3.4. Stop the device

To stop the device:

1. Close the Audiowin® 20 programme window.
2. If the device is not to be used for an extended time, unplug the USB lead from the headset and store it in its packaging.

## 4. Use

## 4.1. Using the hygienic single-use earpad covers

The FIM MEDICAL company has specifically developed hygienic single-use earpad covers to be used with the AUDIOLYSER® ADL20 audiometer.

**Caution:** These hygienic single-use earpad covers should be systematically used for each test, and changed after each patient.

**Caution:** For hygiene and biocompatibility reasons, it is compulsory to use the FIM MEDICAL brand of single-use earpad covers with these audiometric headphones.

These earpad covers have been developed specifically to meet ISO 10993 material biocompatibility requirements, as well as guaranteeing perfect sound transmission whilst respecting IEC 60645-1.

### Using the single-use earpad covers:

- ✓ Place a single-use earpad cover over each of the audiometric headphone earpads (speaker side)
- ✓ Adjust the earpad covers so that there are no material folds between the headphone and the patient

## 4.2. Preliminary patient explanation

### 4.2.1. Headset position

Seat the patient comfortably and help them position the headset over the ears correctly. The pads should be centred over the ear holes, with the headband resting on top of the head, without forcing it.

Red earphone  
on the right ear



Blue earphone  
on the left ear

### 4.2.2. MVA (multilingual vocal assistant)

Using software commands the MVA (multilingual vocal assistant) sends vocal instructions to the patient via the headset. These instructions come in several languages (optional), thus making audiometry screening accessible to a wide range of the public.

The MVA explains the test procedure as well as using the patient response button. It will also intervene if there is a usage error, during the familiarization phase.


#### 4.2.3. *Most frequent errors*

1. Continuous press on the response button.
2. Repeated and untimely presses on the response button.
3. Pressing too lightly.

If the patient is unable to get used to the response button, the perception of sound can be confirmed by raising an arm. In this case, the operator validates the threshold by clicking “Validate”, or pressing “Enter”.

## 4.3. Software use

### 4.3.1. Presentation



### 4.3.2. Checking function

When the Audiowin<sup>®</sup>20 interface is run, the presence indicator turns from grey to blue. If this does not happen, check that:


1. The USB lead is correctly inserted.
2. The software has not been run twice.
3. Installation of drivers was correctly done.

As soon as the presence indicator colours, click on the graph of the left or right ear at the intersection of an audible frequency and intensity.

Transmit the sound by pressing on the keyboard **space** bar; the chosen sound should be perceived.

### 4.3.3. Tool bar

To access main functions.



### 4.3.4. Menu bar

#### 4.3.4.1. File

Function	Description
New	Create a new file
Open	Open a patient file and/or test
Record	Record the current file and test
Export	Export the test performed to the computer
Compare	Compare several curves from the same patient recorded in the data base
Follow up	Follow a patient on an index or frequency over time
Identification	Input or consultation of patient identification
Print	Print the test in progress
Selective print	Print a selection of the test according to certain criteria
Print preview	Display entire pages
Print set up	Choose and set up the printer
Exit	Exit the programme

#### 4.3.4.2. Tests

Function	Description
Automatism	Start/stop automatism
Pause	Automatism pause, or rerun
MVA repeat	Repeat last MVA message to the patient
MVA	List of available languages
Comment	Post a comment related to the test
Results	Display audiometry result and the calculated indexes
+90d Authorization	Tick this menu to go over 90 dB value

#### 4.3.4.3. Options

Function	Description
Parametres	Parameter the software

#### 4.3.4.4. Help

Function	Description
About	Version and copyright information
System information	Environment and device information

## 4.4. Patient identification

### 4.4.1. Identification

Patient identification is required to record test results in the data base, but also to calculate and display the predicted response limits according to the age and gender of the patient.

Click "**Identification**" button before or after the audiometry.

Identification

Last name : \*  ... Company :

First name : \*  Position :

Birth date : \*  Age : 0 Exposure :

Id : \*  Notes :

Gender : \*  Male  Female Operator :  ...

\* = required field

OK Cancel


Fill in the required fields (fields with a red asterisk are compulsory). Go from one field to another by clicking on the box, or by pressing the "**Tab**" button.

Use the small 3-points button (...) located on the right of the last name to check if the patient has already been recorded, using only the first letters of the last name.

Date of birth can be input with or without spaces. The year can be 2 or 4 numbers, e.g.: 21/04/1981 or 21/04/81.

#### 4.4.2. Display of predicted values

After identification, the graph appears as such:



Calculation of predicted values is based on the ISO 7029 standard.


The colour can be changed in the "Parametres" window, "General" tab.

### 4.5. Open a patient file/test

#### 4.5.1. Open a file

To locate the data of a recorded patient, proceed as follows:

Click "Open". The following screen appears:



Enter the first letters of the patient's surname in the "By name" box, then click "Search". The list of names starting with these letters appears. Then click "New exam". A patient search can also be carried out by identification number or date of recording.


#### 4.5.2. Open a test

To visualize or print an old test, the operator can call up a file.

Proceed as seen before. A small + sign is displayed next to the patient name, indicating that tests were done. Click on this + sign.

A list of dates and times are shown.

Double click on one of these dates or click "Load exam" to display the audiogram performed on this date.



## 4.6. Parameter the software

### 4.6.1. Screening

Recommended settings to perform a screening test:

"General" tab.

1. Amplitude variation 5 dB
2. Sound transmission Pulsed

"Automatism" tab.

1. Frequency 1000 Hz, 2000 Hz, 4000 Hz, 6000 Hz, 8000 Hz, 500 Hz, 250 Hz.
2. Lowest hearing level 0 dB
3. Presentation level 50 dB

## "Automatism" tab.

1. Number of tests 3 successive tests achieved
2. Tolerance of doubtful test 10 dB
3. Duration per level Choose a range between 15 - 30 tenths of a second.

All these settings are programmed only once, but can be changed at any time.

### 4.6.2. Window parameters

To define software parameters:



Click on the parameters icon in the tool bar or select **Option** in the menu bar, then **Parameters**.

#### 4.6.2.1. General tab

Function of transmission button on main interface. Sound is automatically transmitted when the mouse cursor is pointed over the button.

Modify the data base path. **Caution: Do not modify this parameter without knowledge of the consequences.**

Display, printing and colour of predicted value.

Action of high and low keyboard arrows.

Modify Audiowin®20 language.

Type of sound transmitted: pulsed or continuous.

Audiowin®20 can print the audiogram in colour (if printer is a colour printer) and choose the logo on the printing sheet.

Date format.

Inhibition of patient response button. The patient signals sound perception by raising one hand. The operator validates the measurement with the Enter key or the Validate button.

Variation of sound amplitude.

Parametering of right and left arrows function on an audiogram. Each frequency is swept in order, or only the selected sequence.

Display categorization

None     Merluzzi 1979  
 Degrees of hearing loss

Changing the display areas categorization of audiograms.

#### 4.6.2.2. Automatism tab

Use this tab to:

- know the criteria used in the sequence by default
- create and manage sequences for automatic mode (refer § 4.9 automatic mode)

Parameters

General Automatism Calculations Export User

Trial

Trials number : 5

Doubtful test tolerance (dB) : 10

Emitting time by level (1/10s)

Fixed 30

Random Min : 15 Max : 20

Sequence

Sequence	
New	Rename
1: 1000	7: 500
2: 1500	8: 250
3: 2000	9: NO
4: 3000	10: NO
5: 4000	11: NO
6: 750	12: NO

Enter the order the frequencies will appear. "NO" indicates that the frequency will not be tested.

First ear :  Right  Left

Alert at the end of test

Low level : 0 dB

Hearing level : 50 dB

MVA Level : 200  MVA Options


Play MVA

First test  Every tests

OK Annuler Aide

#### 4.6.2.1. Onglet Calculs

Calculations MP42 (Legal Index Table No. 42 Occupational Diseases of the general scheme), MAP (Mean Hearing Loss) and IPA (indicator of Precocious Alerts), or DP42.01 (Doenças profissionais 42.01), or HSE<sup>1</sup> categorization, Merluzzi 1979<sup>2</sup> categorization or MPB 2002<sup>3</sup> categorization are submitted by Audiowin<sup>®</sup> 20.




<sup>1</sup> For more information refer to "The Control of Noise at Work Regulations 2005":

<http://www.legislation.gov.uk/uksi/2005/1643/contents/made>

<sup>2</sup> For more information, please consult the "Prevenzione dei danni uditivi rumore da di lavoro in ambient - linee guided proposte dalla società italiana di medicina del lavoro e igiene industrial"

#### 4.6.2.2. Export tab

Use this tab to parameter the destination of the exported file when pressing on the F12 key. Choose the name of the file as well as the format by default.



You can also parameter the export in a PDF file automatically after each recording.

Destination file can be chosen from the files generated, as well as the file name. File name can be customized and certain variables can be used. However, at least two variables cannot be modified: test time and date, so as to differentiate files.

NAME] → Patient surname

[FIRSTNAME] → Patient first name

[ID] → Patient ID

Note: these variables can be automatically filled in by ticking the appropriate box.

#### 4.6.2.3. User tab

Use this window to input your details and display them at the top of the reports page.

The image shows a software dialog box titled "Parameters" with a close button (X) in the top right corner. It has five tabs: "General", "Automatism", "Calculations", "Export", and "User". The "User" tab is selected. The dialog contains the following fields and controls:

- "Last name first name" :
- "Speciality" :
- "Address" :
- "Town code country" :
- "Phone" :
- "Fax" :
- "Print" :

At the bottom right, there are three buttons: "OK", "Annuler", and "Aide".

## 4.7. Display areas categorization

Audiowin20<sup>®</sup> allows the display area categorization to help the operator to get a quick overview of the trend of the results of the current review. Two types of zones are proposed:

**Merluzzi 1979 :**



**Degrees of hearing loss:**



Degrees of hearing loss	Hearing threshold	Hearing ability
None	0 to 20 dB	Considered normal hearing.
Light	21 to 40 dB	Difficult perception of light speaking and conversations, especially in noisy environments. Good perception in a quiet environment.
Moderate	41 to 55 dB	Difficult perception of speaking, especially when background noise. Tend to increase the volume of the television or radio.
Moderate to severe	56 to 70 dB	Speech perception greatly reduced. Participation in group discussions very difficult.
Severe	71 to 90 dB	Inability to hear speech to normal and also trouble with loud noises. The amplification is required.
Deep	91 dB and +	Environmental sounds and speech are almost imperceptible.

*Audiometric classification of hearing impairment based on the recommendations of the International Bureau Audiophonology.*

[http://www.biap.org/index.php?option=com\\_content&view=article&id=5%3Arecommandation-biap-021-bis&catid=65%3Act-2-classification-des-surdites&Itemid=19&lang=fr](http://www.biap.org/index.php?option=com_content&view=article&id=5%3Arecommandation-biap-021-bis&catid=65%3Act-2-classification-des-surdites&Itemid=19&lang=fr)

## 4.8. Manual mode

### 4.8.1. Operation

Audiowin®20 is designed to perform audiometries in automatic and manual mode.

For manual mode, use the keyboard and/or the mouse to:

- Select frequency
- Select intensity
- Transmit sound
- Validate patient response (if the patient response button is not used)

These points are explained in the following chapter.




3-button mouse

Operation	Keyboard	Mouse left click	Mouse scrolling wheel
Sound transmission			Press on scrolling wheel
Select intensity			Use scrolling wheel to modify intensities
Select frequency			Use Ctrl + scrolling wheel to modify frequencies
Validate patient response			

Setting of frequencies and intensities can also be accessed with the mouse, using the right or left audiogram graphs, and clicking on the intersection of a frequency and intensity.

Note: To avoid handling errors, sudden changes in amplitude can be automatically controlled by the software, so as not to damage the patient's ear.

#### 4.8.2. Keyboard keys

		Page UP Page Down	Intensity to value of lowest hearing level. Intensity to a pre-determined high value in the “sequence” tab (presentation level). So as to avoid repeated presses on the intensity settings arrows, use the “ <b>page up</b> ” and “ <b>page down</b> ” keys to pass from one extreme to the other on the intensities scale.
		END	Stop automatic procedure.
		Tab	Go from one ear to the other.
		Enter	Operator validates patient response.
		Space	Transmission of sound in manual mode.
		Del/Erase	Delete a validation on an audiogram.
		F12	Export the audiogram in progress to a file.
		F4	Display audiogram results.

## 4.9. Automatic mode

### 4.9.1. Parameters

To use the audiometer in automatic mode, firstly ensure the programme is set up according to your usual working method.

Possible settings choices:


1. Frequencies tested.
2. Order of frequencies tested.
3. Amplitude variation (steps of 1.5 or 10 dB).
4. Signal type (continuous or pulsed).
5. First ear tested.
6. Number of test(s) per frequency.
7. Response level differences considered doubtful by Audiowin®20.
8. MVA language and volume.

These settings will apply throughout the test, the length of which is inversely proportional to the quality of the final audiogram.

To access the parametering window, in the "Options" menu, choose "Parameters".

Use the tabs located at the top of this window to access different settings.


Click on the "Automatism" tab.



#### 4.9.2. Creation of sequences

Audiowin®20 also offers the possibility of creating personalized sequences: click “New”, then name this sequence. Click “OK”.

Select the frequencies to be tested and the different criteria of the test. In this example, the test will start at 1000 Hz frequency then 2000Hz, 4000Hz and 500 Hz. The other frequencies are positioned to NO.



The new sequence will automatically be recorded in the scrolling menu of the existing sequences.



### 4.9.3. Run automatic test

Place the headset on the patient's head and click "Auto".

The MVA gives the operating instructions to the patient. The test then starts with the familiarization phase at 50 db. During this phase, the software detects abnormalities linked to incorrect presses of the patient response button and the MVA informs the patient via vocal messages such as "Release the button".

The patient should press on the blue part of the patient response button if the sound is perceived. If not, the sound increases in intensity to 90 dB, then displays "No response".

If the patient hears, Audiowin®20 runs the familiarization threshold search at 1000 Hertz, starting at 0 dB.

The programme then carries out the test of all frequencies on the two ears. If there is a doubtful response, a question mark is placed next to the patient response and will return to this test at the end of the sequence. If a doubt persists, a frequency may be retested in manual mode.

Tests in progress can be interrupted temporarily by pressing on the "Pause".

## 4.10. Display of results

Audiowin®20 displays a table of results summarizing the test in progress. It indicates certain other calculations.

Frequency (Hz)	Right (dB HTL)	Left (dB HTL)	Predicted (ISO7029)
125			16
250	0	5	15
500	10	15	16
750	10	15	
1000	10	10	17
1500	10	15	21
2000	15	20	25
3000	10	20	36
4000	20	25	47
6000	20	25	53
8000			63

? = Doubtful result    0 = No response

	Right ear	Left ear	Average		Right ear	Left ear
MP42	13.75	17.50	15.63	Merluzzi 1979 cat.:	0	0
MHL	17.50	22.50	20.00	MPB 2002 cat.:	1a	1a
EWI	16.67	23.33	20.00			

DP42.01	12.50	15.50	14.00
1+2+3+4+6kHz	75	100	
3+4+6kHz	50	70	
1+2+3+4+6kHz	55	75	

HSE cat.: 4 - Rapid hearing loss

Method : Manual

OK

Summary contains:

- Patient results
- MP42 (occupational diseases)
- PAM (Average Hearing Loss)
- IPA (Early Warning Indicator)
- DP42.01 (Doenças profissionais 42.01)
- HSE (Health and Safety Executive) categorization
- Merluzzi 1979 categorization


- MPB 2002 categorization
- Method of test performance

## 4.11. Print results


### 4.11.1. Simple printing

When the audiometric test is completed, and after recording the results, press "**Print**". Printing is immediate.

Printing can also be done from the "File" menu. In this case the following window, depending on printer type, appears:



The document will be printed as such:




Patient identification file

ADL-20 serial number + software version N°

#### 4.11.2. Selective printing

Printing one, or several documents, of tests performed on different dates can be programmed.


Click on the "File" menu then "Selective printing". The following window appears:




If printing several documents, ensure the paper supply is sufficient.

#### 4.11.3. PDF printing


When the test is complete, flick on **File** → **Export** → **PDF**.



Choose the location of the PDF file and the file name, then click on **“Record”**.



If PDF creation is successful, the following message appears:



## 5. Maintenance

## 5.1. Cleaning

---

The device should be cleaned after each use with a damp cloth and a bactericide solution. FIM MEDICAL recommends the use of Bactinyl® odourless cloths.

We recommend asking the patient to disinfect their hands before handling the device.

After each patient, clean the parts of the device in contact with the patient:

- Headset pads (take care not to wet the headphones)
- Head band
- Patient response button

The use of spray is not recommended; a misdirected jet may definitively damage the headphones.

## 5.2. Maintenance

---

To conform with the ISO 8253-1:2010 standard, audiometric equipment should be regularly checked and calibrated. Recommended:

- **Daily:** clean, check general state of equipment, check the device functions over the entire frequency range, check patient response button.  
Checks should be made in the same environment as patient tests.
- **Annual service** performed by the manufacturer.

Only FIM MEDICAL is qualified to perform annual servicing of audiometers.

A calibration certificate is delivered.

The expiry date of the FIM MEDICAL calibration is indicated by the software. Before this expiry date, please return the device to FIM MEDICAL for calibration.

**Caution: Have the device recalibrated before the expiry date. After expiry date, a drift may falsify results.**

## 5.3. Guarantee

---

Within the framework of the contractual guarantee of one year, only repairs are covered. The guarantee is only applicable if normal and usual conditions of use are respected. During annual servicing, a certain number of preventive operations are performed; breakdowns following annual service are not covered by the guarantee.

## 6. Available Accessories

## 6.1. Hygienic single-use earpad covers

---

FIM MEDICAL has specifically developed single-use earpad covers for use with the AUDIOLYSER® ADL20 audiometer.

**Warning:** These single-use earpad covers must be used systematically for each test and replaced between each patient.

**Warning:** For hygiene and biocompatibility reasons, it is essential to use the FIM MEDICAL brand single-use hygienic earpad covers with audiometric headphones.

**These earpad covers have been specifically developed to meet the biocompatibility requirements of ISO 10993 materials and to guarantee perfect sound transmission in accordance with IEC 60645-1.**

If you would like to purchase the FIM MEDICAL single-use hygienic earpad covers, please contact FIM MEDICAL or your distributor. Various packaging options are available.

## 7. FAQ

## 7.1. No sound is perceived

---

- Check the headset is connected.
- Check the indicator cone of headset presence on the software interface is coloured blue.

## 7.2. Error message appears at recording

---

### 7.2.1. *"Identification incomplete"*

- Check all the compulsory identification fields are filled in.

### 7.2.2. *"Operation must use updateable query"*

- This message is due to a problem of writing access rights on that computer. The administrator must give all tree structure rights, where the data base is found, to the operator.

## 7.3. Patient identification file found but not the tests

---

- Click "+" on the left of the patient's name in the "Open" dialogue box.



*Thank you for reading this manual.  
If you require further information please don't  
hesitate to contact us.*



51 rue Antoine Primat 69100 Villeurbanne - FRANCE

Tel: (+33)04 72 34 89 89 - Fax: (+33)04 72 33 43 51

[contact@fim-medical.com](mailto:contact@fim-medical.com) / [www.fim-medical.com](http://www.fim-medical.com)